



The **897 Dosimeter Sound Measuring System** conducts personal noise surveys to meet IEC and OSHA requirements.

Used as a sound level meter, the 897 provides a quick survey of targeted areas and identifies suspect locations. The 897 is also ideal for industrial and environmental sound level measurements and checking noise areas for excessive dB exposure.

Use the 897 to take measurements and record, then view and print reports from the computer by using HyperTerminal program for later use during OSHA or other official inspections.

NOTE: Microsoft no longer includes a HyperTerminal program on Windows 7.

The 897 computes dosage and records up to 31 hours, at one minute intervals, of Lavg and Lmax readings and the number of detected 140 dB peak occurrences per minute. You can then print data up to 8 hours of recorded readings.

- Conforms to ANSI S1.4-1983, ANSI S1.25-1978, and IEC 651-1979 Specifications
- Choice of exchange rates (3, 4, or 5 dB) and criterion levels (70, 80, 84, 85, or 90 dB)
- Selectable threshold Level from 50-99 dB in 1 dB Increments
- Dual range (50-100 dB and 80-130 dB)
- Real Time Clock with rechargeable battery
- Security Lockout feature with Internal Data Storage
- Self-Test Modes test operational readiness

Rev. 04-15



Ordering Information	
Sound Dosimetry System	Catalog Number
897 Universal Noise Dosimeter	12645
SMS-2 Universal Noise Dosimeter Kit	12646
887-2 Calibrator	12648
Accessories	Catalog Number
RS-232 Serial Cable for 897	02233
Parallel Printer Cable for 897 (Optional)	02234
Case, Molded Plastic	45021

SMS-2 Universal Sound Dosimeter Kit

The SMS-2 contains everything necessary to perform accurate surveys for OSHA and IEC requirements, along with the performance and reliability Simpson's test instruments have to offer.

Kit In	cludes:
Α	Rugged Carrying Case
В	AC Adaptor (110, 120, 220, 240)
С	RS-232 Serial Cable and (Optional) Parallel Printer Cable
D	Multi-Spline Wrench
Е	Calibration Screwdriver
F	Microphone Clip
G	Wind Screen
Н	897 Dosimeter
I	9V Battery
J	887-2 Calibrator
K	Nose Cone



Rev. 04-15

2



Specifications of Mode	1 907					
Specifications of Mode	51 O3 <i>1</i>					
GENERAL						
Power Requirements	Ni-Cad Rechargeable battery pack					
Warm-Up Time	1 minute					
Dimensions	6.8" (H) x 3.1" (W) x 1.1" (D), (172x80x28mm)					
Weight	18 oz. (500g) including battery pack					
Case Construction	Anodized aluminum extrusion					
Security Feature	Locks-in operating function					
Real Time Clock	Clock continues running in all functions					
Performance Capability	Conforms to ANSI S1.4-1983 and IEC 651-1971 for sound level function					
	and ANSI S1.25-1978 for dosimeter function					
SOUND LEVEL MODE						
Ranges	50 to 100dB, 80 to 130 dB					
Accuracy	True RMS, Type S2					
Dynamic Range	Selectable from 50 to 100dB and 80 to 130dB					
Frequency Response Weighting	"A" Weighting					
Response Time	Slow (1 second)					
Crest Factor	10:1 at maximum indication					
Frequency Range	31.5Hz to 8KHz					
DOSIMETER MODE	To 1 4 11 6 To 20 ID 1 4 ID 1					
Threshold Level	Selectable from 50-99dB in 1dB increments					
Criterion Level	Selectable, 70, 80, 84, 85 or 90dB					
Criterion Duration (TC)	Eight (8) Hours					
Exchange Rate	Selectable 3, 4 or 5dB					
Maximum Indication	999.9% DOSE					
Resolution	0.1% DOSE to 999.9%					
Elapsed Time	Displays either Hours: Minutes: Seconds, up to 99 Hours 59 Minutes, 59 Seconds. Timing Accuracy: 0.05%					
140dB Peak	Displays number of 140dB peaks that have been detected					
1100D 1 Out	Displays frames of Freez pound that have been detected					
MICROPHONE						
Туре	0.52" (13.2mm) Diameter Electret Condenser, Omnidirectional 70° angle of incidence.					
••	Maximum sound pressure level 148dB.					
DISPLAY	T					
Numerical	4 Digit LCD					
Annunciators	Lo Batt, Spl Max, Int 60, % Dose, dBA, plus a colon and decimal point					
ENVIRONMENTAL						
Operating Temperature	-10° to 50°C					
Temperature Coefficient	±0.05dB per °C (25° to 50°C) for reference measurement of 105dB at 1000Hz					
Relative Humidity	Range: 0-95% RH					
	Influence: Less than 0.5db over measurement range					
Storage Range	-30° to 45°C (Limitation of battery)					
Magnetic Field Influence	No effect in of 1 oersted (80A/m)					

3

Rev. 04-15



STANDARD REPORT OUTPUT

		SIMPSON 8					
		SOUND AN		PORT			
		Т	YPE 2A				
CRITERION = 90	DR						
THERSHOLD = 8							
EXCHANGE RAT							
UNIT IDENTIFAC	ATION # 25						
100							
JOB							
NAME							
LOCATION							
START DATE: 21							
START TIME: 14: CALIBRATION: 9		1.24 21/EED/05	DANCE, 50	0/100 DP			
CALIBRATION: 9							
CALIBRATION: I	I NOU E.C.	7.24 Z I/FED/95	HANGE: 8	50/ TOO DB			
MEASUREMENT	SUMMARY:	:					
RUN TIME	= 00:15						
HOLD TIME	= 00:02	!					
L EQ	= 87.2	DBA					
SPL MAX	= 87.2						
DOSE	= 1.59	%					
140 DB PEAKS	= 0						
		HI	SOTGRAM				
		HI	SOTGRAM	l			
		60 70	80 90	0 100	110	120	130
HRS:MIN	++	60 70	80 90	0 100	110	120	130
14:25	++	60 70 + + + + .	80 90 .++	0 100	110	120	130
14:25 14: 230	++	60 70 + + + + .	80 90 .++ ==== 88 == 86	0 100	110	120	130
14:25 14: 230 14:35	++	60 70 + + + + . =================	80 90 .++ ==== 88 == 86 HOLD	0 100	110	120	130
14:25 14: 230	++	60 70 + + + + .	80 90 .++ ==== 88 == 86 HOLD	0 100	110	120	130
14:25 14: 230 14:35 HRS:MIN	++ ======= ++	60 70 + + + + ========= 76 + + + +	80 90 .++ ==== 88 == 86 HOLD .++	0 100 + +	RANC	SE.	130
14:25 14: 230 14:35 HRS:MIN HRS:MIN 14:25	++ ======= ++ LAVG 86.9	60 70 + + + + ========= 76 + + + + LMAX 95.6	80 90 .++ ==== 88 == 86 HOLD .++	0 100 + + EAKS	RANC CHANGI	SE.	130
14:25 14: 230 14:35 HRS:MIN HRS:MIN 14:25 14:26	++ ======= ++ LAVG 86.9 93.2	60 70 + + + + ========= 76 + + + + LMAX 95.6 106.9	80 90 .++ ==== 88 == 86 HOLD .++	0 100 + + EAKS	RANC CHANGI 80/130	SE.	130
14:25 14: 230 14:35 HRS:MIN HRS:MIN 14:25 14:26 14:27	++ ======= ++ LAVG 86.9 93.2 87.9	60 70 + + + + ========= 76 + + + + LMAX 95.6 106.9 92.5	80 90 .++ === 88 == 86 HOLD .++ PE 0 0	0 100 + + EAKS	RANC CHANGI 80/130 80/130	SE.	130
14:25 14: 230 14:35 HRS:MIN HRS:MIN 14:25 14:26 14:27 14:28	++ ======= ++ LAVG 86.9 93.2 87.9 87.3	60 70 + + + + ========= 76 + + + + LMAX 95.6 106.9 92.5 97.3	80 90 .++ ==== 88 == 86 HOLD .++ PE 0 0 0 0	0 100 + + EAKS	RANC CHANGI 80/130 80/130 80/130	SE.	130
14:25 14: 230 14:35 HRS:MIN HRS:MIN 14:25 14:26 14:27 14:28 14:29	++ ======= ++ LAVG 86.9 93.2 87.9 87.3 86.4	60 70 + + + + ========= 76 + + + + LMAX 95.6 106.9 92.5 97.3 86.6	80 90 .++ ==== 88 == 86 HOLD .++ PE 0 0 0 0	0 100 + + EAKS	RANC CHANGI 80/130 80/130 80/130 80/130	SE.	130
14:25 14: 230 14:35 HRS:MIN HRS:MIN 14:25 14:26 14:27 14:28 14:29	++ ======= ++ LAVG 86.9 93.2 87.9 87.3	60 70 + + + + ========= 76 + + + + LMAX 95.6 106.9 92.5 97.3	80 90 .++ ==== 88 == 86 HOLD .++ PE 0 0 0 0	0 100 + + EAKS	RANC CHANGI 80/130 80/130 80/130 80/130 80/130	SE.	130
14:25 14: 230 14:35 HRS:MIN HRS:MIN 14:25 14:26 14:27 14:28 14:29	++ ======= ++ LAVG 86.9 93.2 87.9 87.3 86.4 86.5	60 70 + + + + ========= 76 + + + + LMAX 95.6 106.9 92.5 97.3 86.6 93.0	80 90 .++ ==== 88 == 86 HOLD .++ PE 0 0 0 0 0	0 100 + + EAKS	RANC CHANGI 80/130 80/130 80/130 80/130	SE.	130
14:25 14: 230 14:35 HRS:MIN HRS:MIN 14:25 14:26 14:27 14:28 14:29 14:30 14:31	++ ======= ++ LAVG 86.9 93.2 87.9 87.3 86.4 86.5 87.2	60 70 + + + + ========= 76 + + + + LMAX 95.6 106.9 92.5 97.3 86.6 93.0 90.0	80 90 .++ ==== 88 == 86 HOLD .++ PE 0 0 0 0 0 0	0 100 + + EAKS	RANC CHANGI 80/130 80/130 80/130 80/130 80/130	SE.	130
14:25 14: 230 14:35 HRS:MIN HRS:MIN 14:25 14:26 14:27 14:28 14:29 14:30 14:31 14:32	++ ======= ++ LAVG 86.9 93.2 87.9 87.3 86.4 86.5 87.2 87.2	60 70 + + + + ========== 76 + + + + LMAX 95.6 106.9 92.5 97.3 86.6 93.0 90.0 91.2	80 90 .++ ==== 88 == 86 HOLD .++ PE 0 0 0 0 0 0 0	0 100 + + EAKS	RANC CHANGI 80/130 80/130 80/130 80/130 80/130 80/130	SE.	130
14:25 14: 230 14:35 HRS:MIN HRS:MIN 14:25 14:26 14:27 14:28 14:29 14:30 14:31 14:32 14:33	++ ======== ++ LAVG 86.9 93.2 87.9 87.3 86.4 86.5 87.2 87.2 84.9	60 70 + + + + =	80 90 .++ ==== 88 == 86 HOLD .++ PE 0 0 0 0 0 0 0 0 0	0 100 + + EAKS	RANC CHANGI 80/130 80/130 80/130 80/130 80/130 80/130 80/130	GE ED	130
14:25 14: 230 14:35 HRS:MIN HRS:MIN 14:25 14:26 14:27 14:28 14:29 14:30 14:31 14:32 14:33 14:34	++ ==============================	60 70 + + + + =	80 90 .++ === 88 == 86 HOLD .++ PE 0 0 0 0 0 0 0 0	0 100 + + EAKS	RANC CHANGI 80/130 80/130 80/130 80/130 80/130 80/130 80/130	GE ED	130 HOLD
14:25 14: 230 14:35 HRS:MIN HRS:MIN 14:25 14:26 14:27 14:28 14:29 14:30 14:31 14:32 14:33 14:33 14:34 14:35 14:36 14:37	++ ======== ++ LAVG 86.9 93.2 87.9 87.3 86.4 86.5 87.2 87.2 84.9 85.4 86.2 86.5 60.9	60 70 + + + + = = = 76 + + + + + + + + .	80 90 .++ === 88 == 86 HOLD .++ PE 0 0 0 0 0 0 0 0 0 0 0 0 0	0 100 + + EAKS	RANC CHANGI 80/130 80/130 80/130 80/130 80/130 80/130 80/130 CHANGI 50/100 50/100	GE ED	
14:25 14: 230 14:35 HRS:MIN HRS:MIN 14:25 14:26 14:27 14:28 14:29 14:30 14:31 14:32 14:33 14:34 14:35 14:35 14:36 14:37 14:38	++ ===================================	60 70 + + + + ===================	80 90 .++ ==== 88 == 86 HOLD .++ PE 0 0 0 0 0 0 0 0 0 0 0 0 0	0 100 + + EAKS	RANC CHANGI 80/130 80/130 80/130 80/130 80/130 80/130 80/130 CHANGI 50/100 50/100	GE ED	HOLD
14:25 14: 230 14:35 HRS:MIN HRS:MIN 14:25 14:26 14:27 14:28 14:29 14:30 14:31 14:32 14:33 14:33 14:34 14:35 14:36 14:37	++ ======== ++ LAVG 86.9 93.2 87.9 87.3 86.4 86.5 87.2 87.2 84.9 85.4 86.2 86.5 60.9	60 70 + + + + = = = 76 + + + + + + + + .	80 90 .++ === 88 == 86 HOLD .++ PE 0 0 0 0 0 0 0 0 0 0 0 0 0	0 100 + + EAKS	RANC CHANGI 80/130 80/130 80/130 80/130 80/130 80/130 80/130 CHANGI 50/100 50/100	GE ED	HOLD
14:25 14: 230 14:35 HRS:MIN HRS:MIN 14:25 14:26 14:27 14:28 14:29 14:30 14:31 14:32 14:33 14:34 14:35 14:35 14:36 14:37 14:38	++ ===================================	60 70 + + + + ===================	80 90 .++ ==== 88 == 86 HOLD .++ PE 0 0 0 0 0 0 0 0 0 0 0 0 0	0 100 + + EAKS	RANC CHANGI 80/130 80/130 80/130 80/130 80/130 80/130 80/130 CHANGI 50/100 50/100	GE ED	HOLD



887-2 Calibrator

The model 887-2 calibrator has a selectable sound pressure range of 94dB and 114dB at 1 KHz.

The calibrator is immune to a wide range of temperature and humidity conditions is compatible with 1/2" microphone.

Operational readiness is indicated by the low battery red LED.



GENERAL				
Battery	One 9-volt battery, NEDA 1604A			
Battery Life	Approximately 35 hours for 2 hours per day operation, with 9V alkaline batter			
Mechanical Construction	Aluminum case includes acoustic cavity and provides shielding			
Weight	13.5oz (.35kg) - 897, 5.9 lbs. (2.68 Kg) SMS-2 Kit			
Dimensions	5 1/4" long, 2" diameter (131 mm long, 50mm diameter)			
ACOUSTIC OUTPUT				
Microphone	0.52" inch diameter			
Frequency	1000Hz ± 1%			
Sound Pressure	Selectable at 94 or 114 dB			
Reference Level	0dB = 20μ Pascals			
Accuracy	± 0.5dB at reference conditions			
Distortion	Less than 2%			
ENVIRONMENTAL				
Operating Temperature	0° to 50°C			
Storage Temperature	-40 to +60°C, battery removed			
Temperature Coefficient	Less than ± 0.05 dB/°C			
Operating Relative Humidity	0 to 90%, non-condensing			
Reference Conditions	23°C, 760mm Hg			
Relative Humidity	30 to 60%			

5

Rev. 04-15